



*Up-to-date Questions and Answers from authentic resources to improve knowledge and pass the exam at very first attempt. ----- Guaranteed.*



RPFT Dumps  
RPFT Braindumps  
RPFT Real Questions  
RPFT Practice Test  
RPFT Actual Questions



**Medical**

**RPFT**

*Registered Pulmonary Function Technologist*



<https://killexams.com/pass4sure/exam-detail/RPFT>

Question: 102

In setting up a CO analyzer, a pulmonary function technologist notices that the analyzer reads -0.03 while sampling air. The technologist should

- A. Accept the reading because it is within  $\pm 3\%$ .
- B. Adjust the reading to +0.03.
- C. Adjust the reading to 0.00.
- D. Reverse the sample flow.

Answer: C

Question: 103

A patient's vital capacity is slightly reduced, the FEWFVC is normal, and the uncorrected DLco is increased. Which of the following is the most likely diagnosis?

- A. diffuse pulmonary fibrosis
- B. diaphragmatic hemiparesis
- C. kyphoscoliosis
- D. polycythemia vera

Answer: D

Question: 104

The following results are obtained from an adult male:

<b>Uncorrected <math>D_{LCO}</math></b>	<b>32 mL/min/mm Hg (STPD)</b>
<b>Hb</b>	<b>14.6 gm/dL</b>
<b>COHb</b>	<b>1.2%</b>
<b>Alveolar volume</b>	<b>3500 mL</b>

The corrected DLco value -

- A. is unchanged.
- B. is higher.
- C. is lower.
- D. cannot be calculated.

Answer: A

Question: 105

To check the reliability of a pulse oximeter reading, a pulmonary function technologist should

- A. Calculate the SaO<sub>2</sub> from pH and PaO<sub>2</sub>
- B. Perform hemoximetry
- C. Measure the hematocrit
- D. Have the patient hyperventilate

Answer: B

Question: 106

A 54-year-old male who smokes presents to the pulmonary laboratory for chronic cough and dyspnea on exertion. PFT and blood gas results show the following:

<b>FVC</b>	<b>3.0 L</b>
<b>FEV<sub>1</sub></b>	<b>1.56 L</b>
<b>FEV<sub>1</sub>/FVC</b>	<b>52%</b>

<b>pH</b>	<b>7.39</b>
<b>PaCO<sub>2</sub></b>	<b>45 torr</b>
<b>PaO<sub>2</sub></b>	<b>68 torr</b>
<b>HCO<sub>3</sub><sup>-</sup></b>	<b>28 mEq/L</b>
<b>SaO<sub>2</sub></b>	<b>86%</b>
<b>COHb</b>	<b>8%</b>

Which of the following should the pulmonary function technologist recommend?

- A. DLco measurement
- B. Oxygen therapy with exercise
- C. Trial of varenicline (Chantix)
- D. Lung volume measurement

Answer: A

Question: 107

Which of the following is a suitable policy for following Standard Precautions in a pulmonary function laboratory?

- A. Eye protection is required when obtaining ABGs from patients with hepatitis.
- B. Reusable mouthpieces should be disposed when a patient has a history of tuberculosis.
- C. Gloves are optional when obtaining arterial blood samples using a kit
- D. Reusable mouthpieces should be disinfected between each patient.

Answer: B

Question: 108

While assessing a patient's expired gases at rest prior to exercise, a pulmonary function technologist calculates the RER as 0.6. Which of the following is the most likely explanation?

- A. B. gas analyzer is malfunctioning
- B. A gas analyzer is malfunctioning
- C. The expired gas is contaminated with air
- D. The patient is hyperventilating

Answer: B

Question: 109

Which of the following is an appropriate reason to perform a multiple-breath nitrogen washout test?

- A. Measure anatomical dead space.
- B. Differentiate obstruction from restriction.
- C. Detect early small airway disease.
- D. Measure oxygen consumption.

Answer: C

Question: 110

During daily quality control procedures on an infrared CO<sub>2</sub> analyzer, a pulmonary function technologist is unable to adjust the gain to the calibration gas concentration. Which of the following is the most likely explanation?

- A. Water droplets in the sample cell
- B. Saturation of the soda lime
- C. Presence of high levels of oxygen
- D. Increased gas sampling rate

Answer: A

Question: 111

A patient who is about to begin pulmonary function testing is visibly upset and complains to a pulmonary function technologist that she felt a receptionist was rude to her. Which of the following should the technologist do?

- A. Try to get the patient to calm down by telling her that the receptionist is probably just having a bad day.
- B. Give the patient time to calm down and ask the laboratory manager to become involved.
- C. Ignore the complaint because it is not going to affect the testing about to begin.
- D. Accompany the patient back to the reception area and try to determine who was at fault.

Answer: B



# SAMPLE QUESTIONS

*These questions are for demo purpose only. **Full version** is up to date and contains actual questions and answers.*

*Killexams.com is an online platform that offers a wide range of services related to certification exam preparation. The platform provides actual questions, exam dumps, and practice tests to help individuals prepare for various certification exams with confidence. Here are some key features and services offered by Killexams.com:*

**Actual Exam Questions:** *Killexams.com provides actual exam questions that are experienced in test centers. These questions are updated regularly to ensure they are up-to-date and relevant to the latest exam syllabus. By studying these actual questions, candidates can familiarize themselves with the content and format of the real exam.*

**Exam Dumps:** *Killexams.com offers exam dumps in PDF format. These dumps contain a comprehensive collection of questions and answers that cover the exam topics. By using these dumps, candidates can enhance their knowledge and improve their chances of success in the certification exam.*

**Practice Tests:** *Killexams.com provides practice tests through their desktop VCE exam simulator and online test engine. These practice tests simulate the real exam environment and help candidates assess their readiness for the actual exam. The practice tests cover a wide range of questions and enable candidates to identify their strengths and weaknesses.*

**Guaranteed Success:** *Killexams.com offers a success guarantee with their exam dumps. They claim that by using their materials, candidates will pass their exams on the first attempt or they will refund the purchase price. This guarantee provides assurance and confidence to individuals preparing for certification exams.*

**Updated Content:** *Killexams.com regularly updates its question bank and exam dumps to ensure that they are current and reflect the latest changes in the exam syllabus. This helps candidates stay up-to-date with the exam content and increases their chances of success.*

**Technical Support:** *Killexams.com provides free 24x7 technical support to assist candidates with any queries or issues they may encounter while using their services. Their certified experts are available to provide guidance and help candidates throughout their exam preparation journey.*

For More exams visit <https://killexams.com/vendors-exam-list>  
Kill your exam at First Attempt....Guaranteed!